

# Aer 5400



## Trolley or wall mounted Alpha/Beta Continuous Air Monitor



## **Applications:**

- for monitoring airborne radioactive substance (LLRD) activity concentrations and measuring radon / thoron equivalent equilibrium concentration (EECRn & EECTh) and/or potential alpha energy concentration (PAEC) at workplaces
- in nuclear facilities
- in the NORM industry
- in mining operations
- in nuclear medicine (Th-227, Ra-223 and Rn-219)
- For taking samples from ducts and chimneys of nuclear facilities

#### Features:

- continuous monitoring of breathing air for airborne long-lived radioactive aerosols (LLRD) and short-lived radon decay products
- assessment and minimization of inhalation hazards for workers
- alert workers to high levels of airborne activity
- spectroscopic separation of the nuclides and complete compensation of the natural radon background for the LLRD measurement
- menu navigation via touch screen
- all parameters relevant for reliable operation are continuously monitored and are part of the stored measurement data
- flexible alarm system
- filter tape for 330 steps or nearly 1 year at one filter change/day







**Detector type** 400 or 1200mm<sup>2</sup> ion-implanted silicon detector

option "G": double detector for dynamic gamma background

compensation (2 x 1200 mm2 each detector)

**Energy range** 80-150keV...3MeV (beta) (depends on detector size)

3...10MeV (alpha);

**Counting efficiency** approx.. 20% ( $4\pi$ )

Sampling open face and closed sampling with automatic filter sealing

for minimum collection losses

Filter /Stepper membrane-filter tape (PTFE), 5µm pore size; length 30m,

width: 65 mm, sufficient for >330 Filter changes

deposition rate >99.9%

filter pressure mechanism for maximum tightness of the col-

lecting device

active filter test with respect to perforation and exhaustion

fast, tool-less replacement of filter coils

more than 12 month autonomous operation in "normal" envi-

ronment

configurable trigger for filter stepping (e.g. each sample interval, fixed period, filter exhaustion, activity detected)

required period for filter replacement <2s

**Pump** built-in or extern ("Rn"- version)

oil-less, long-life, low noise quality rotary van pump (Becker)

(extern) OR membrane type pump (intern)

nominal air flow 35 SLPM (adjustable range 8 to 50 SLPM

depending on pump type and instrument version)

processor controlled air flow for constant deposition condi-

tions (mass flow sensor)

pressure drop across the filter 15...100mbar (at 35 SLPM)

noise emission approx. 60dBA (in 1m distance)

other vacuum supplies may be connected instead of the

pump

Weight of ext.pumps

VT4.4 – 7 kg (4,1 m<sup>3</sup>/h) VT4.8 – 11,5 kg (8,0 m<sup>3</sup>/h)





#### **Results** exposure for alpha and beta emitters (LLRD) in Bqh/m<sup>3</sup>

dose for Alpha and Beta emitters in  $\mu Sv$  or DAC-hrs (dose

coefficients adjustable by user)

detection of Natural Uranium with automatic selection of the

U<sub>nat</sub> dose coefficient

average activity concentration for Alpha and Beta emitters in

Bq/m³

(E)quilibrium (E)quivalent (C)oncentration (EEC) for radon

and thoron daughter products in Bq/m³

(P)otential (A)lpha (E)nergy (C)oncentration (PAEC) for ra-

don and thoron daughter products in nJ/m<sup>3</sup>

separate channel for alpha gross counting in cps or Bq

option: dose rate in µSv/h

option: gamma spectrum (for Nal detector)

option: gas concentration in ppm

temperature, humidity, pressure, battery voltage

flow rate, filter exhaustion, filter stepping, end of filter tape

signals Alert, Warning, Good

Standards IEC 60761-1

IEC 60761-2 IEC 61578 IEC 61577-3 IEC 1263 CE,VDE

DIN ISO 16639 (VDE 0493-1-6639)

### **Compensation** compensation of natural Radon background by Alpha spec-

troscopy with dynamic fitting of peak shape with respect to

progressive filter exhaustion

upper alpha energy threshold for LLRD = 5,6MeV

static compensation of gamma background

option: dynamic compensation of gamma background by

double detector

dynamic shock rejection (mechanical shock) by pulse signal

shape analysis









**LLRD Sensitivity** approx. 25 cpm/(Bqh/m³)

**Messbereich** 0 ... 10.000 Bqh/m³ (0 ... 50 000 DACh(Pu))

0,6 MBq/m<sup>3</sup> over 1 minute

**Measurement** up to 16 user definable sampling cycles (1s to 1year)

predefined sampling cycles 1, 5, 15, 60 minutes

predefined test cycles

**Detection limits** see tables below

**Alert indication** configurable alert thresholds for all measured results

alert tower with green, yellow and red light, 360° visible

90dB signal buzzer

alert indication at display

alert reset is configurable (either with confirmation by the user or automatic reset if the alert condition is no longer pre-

sent)

pre-defined alerts for LLRD activity, low/high count rate, filter

perforation, end of filter tape

Data storage 2 GB SD-card (> 800 000 data records)

storage of all measured raw data incl. spectra

**Operation / Display** touch screen 6cm x 9cm (4.5"); Graphic 240 x128

high contrast even in direct sunlight

backlight key switch

intuitive, straight forward menu structure

Interfaces USB, RS232, RS422/RS485

option: Net Monitors wireless (ZigBee)

option: TCP/IP (Ethernet/WLAN)

6 additional configurable analogous sensor inputs

1 additional counter input (for models without GM-tube op-

tion only)

3 alert outputs related to the signal lights

1 additional switch contact for external components

**Power supply** 230 VAC/50 Hz (option 110 VAC/60 Hz)

approx. 500 VA

internal NiMH-buffer battery 12 V / 3,8 Ah for more than 12 h operation in case of mains power interruption (without pump)

ATEX category no

**Housing** stainless steel IP65

ease for decontamination

**Versions** wall mounted or trolley







**Dimensions** 540mm x 360mm x 200 mm

<18 kg (wall mounted cabinet) 1350mm x 510mm x 360mm <35 kg (with trolley and pump)

## **Environmental conditions**

Temperature 0 ... 50 °C

Rel. humidity 5 ... 95 % rF. non-condensing

Bar. pressure 800 ... 1100 mbar

Software dVISION remote control

data transfer, visualization

data management, export to text files

system configuration

creating / editing of measurement cycles

network management

**Additional options** 

(on request)

separate filter unit (connection by hose and cable)

GM counter for dose rate measurement

GPS receiver

electrical valve for flow regulation (wall-mounted version, for working with a vacuum supply provided by the customer)

**Calibration /Test** factory calibration in a radon daughter product atmosphere

with aerosol generator

test sources Am-241 (Alpha) and Cs-137 (Beta); recommended are area sources with 25mm or 36mm diameter and

185Bq nominal activity such as Eckert & Ziegler AMRB22757/CDRB22758 (d 30 mm x 0.8 mm)

flow rate check on top of the filter using adapter dome and

low differential pressure air flow meter ( $\Delta p < 15$ mbar

@35l/min)









#### **Detection Limits**

The detection limits stated in the tables below are valid for following operational conditions:

- flow rate = 35l/min
- $k_{1-\alpha} = 3 (99.8\%)$
- $k_{1-\beta} = 1.65 (95\%)$
- $1DAC(Pu) = 0.2Bq/m^3 (10CRF835)$
- 1DAC(Sr90) = 200Bq/m<sup>3</sup> (10CRF835)

Additionally for beta measurement:

- F = 0.6
- gamma background = 0.1µSv/h

The assumption for the detection limit of the concentration is a momentarily step-like increase of air activity concentration up to the detection limit at the beginning of a sampling interval. Furthermore, it is presumed that there was no LLRD activity deposited on the filter.

Alpha LLRD											
Po-218 *)	Detection limit T = 1min			Detection limit T = 5min			Detection limit T = 15min				
Bq/m³	Bqh/m³	DACh	Bq/m³	Bqh/m³	DACh	Bq/m³	Bqh/m³	DACh	Bq/m³		
10	0.92	4.6	55	0.38	1.9	4,6	0.22	1.1	0.9		
20	1.25	6.2	75	0.54	2.7	6,5	0.33	1.6	1.3		
50	1.92	9.6	115	0.88	4.4	10,6	0.58	2.9	2.3		
100	2.70	13.5	168	1.33	6.7	16,0	0.95	4.7	3.8		


Beta LLRD											
Po-218 *)	Detection limit T = 1min			Detection limit T = 5min			Detection limit T = 15min				
Bq/m³	Bqh/m³	DACh	Bq/m³	Bqh/m³	DACh	Bq/m³	Bqh/m³	DACh	Bq/m³		
10	2.75	0.014	165	1.21	0.006	14.5	0.69	0.004	2.8		
20	3.74	0.019	224	1.65	0.008	19.8	0.95	0.005	3.8		
50	5.76	0.029	345	2.55	0.013	30.7	1.47	0.007	5.9		
100	8.06	0.040	483	3.58	0.018	43.0	2.06	0.010	8.3		

<sup>\*)</sup> The activity concentration of Po-218 is always less than the one of Rn-222





#### Possible modifications of Air Monitor Aer 5X00-XXXX



<sup>\*)</sup> options XFG and G cannot be combined



